



In-Plate Preservation

Table of Contents

1	Components	3
1.1	Kit Contents	3
1.2	Components to be supplied by the user	4
2	Step-by-Step guide	5
2.1	Overview.....	5
2.2	Gelation	5
2.3	Recovery.....	6
2.4	Conditioning Chronos Advance CRT container	7
3	Statements	8
3.1	Kit storage and stability.....	8
3.2	Trademarks.....	8
3.3	Cellular material.....	8
3.4	Trademarks.....	8

1 Components

1.1 Kit Contents

Product Code	Components	Units	Unit Volume	Medium to Add
WR-Z096-001	Gel A (5x)	1 tube	2.2 mL	8.8 mL
	Gel B (5x)	1 tube	2.2 mL	8.8 mL
	Gel C (5x)	1 tube	2.2 mL	8.8 mL
	Gelation Buffer 1	1 tube	10 mL	-
	Gelation Buffer 2	1 tube	10 mL	-
	Dissolution Buffer	2 tube	20 mL	-
	Adhesive Plate Seals	2		
WR-Z096-003	Gel A (5x)	3 tubes	2.2 mL	8.8 mL
	Gel B (5x)	3 tubes	2.2 mL	8.8 mL
	Gel C (5x)	3 tubes	2.2 mL	8.8 mL
	Gelation Buffer 1	2 tubes	15 mL	-
	Gelation Buffer 2	2 tubes	15 mL	-
	Dissolution Buffer	6 tubes	20 mL	-
	Adhesive Plate Seals	4	-	-
WR-Z096-006	Gel A (5x)	6 tubes	2.2 mL	8.8 mL
	Gel B (5x)	6 tubes	2.2 mL	8.8 mL
	Gel C (5x)	6 tubes	2.2 mL	8.8 mL
	Gelation Buffer 1	3 tubes	20 mL	-
	Gelation Buffer 2	3 tubes	20 mL	-
	Dissolution Buffer	12 tubes	20 mL	-
	Adhesive Plate Seals	7	-	-
WR-Z096-012	Gel A (5x)	12 tubes	2.2 mL	8.8 mL
	Gel B (5x)	12 tubes	2.2 mL	8.8 mL
	Gel C (5x)	12 tubes	2.2 mL	8.8 mL
	Gelation Buffer 1	1 bottle	120 mL	-
	Gelation Buffer 2	1 bottle	120 mL	-
	Dissolution Buffer	24 tubes	20 mL	-
	Adhesive Plate Seals	13	-	-

Product Code	Components	Units	Unit Volume	Medium to Add
WR-Z096-024	Gel A (5x)	24 tubes	2.2 mL	8.8 mL
	Gel B (5x)	24 tubes	2.2 mL	8.8 mL
	Gel C (5x)	24 tubes	2.2 mL	8.8 mL
	Gelation Buffer 1	2 bottles	120 mL	-
	Gelation Buffer 2	2 bottles	120 mL	-
	Dissolution Buffer	48 tubes	20 mL	-
	Adhesive Plate Seals	25	-	-

NOTE: Remove components from 2-8°C storage for at least 20 minutes before use

1.2 Components to be supplied by the user


96-well plate with adherent cell cultures

1000 µL and 200 µL micropipettes and tips (multichannel pipette is optional)

Cell Culture Medium

2 Step-by-Step guide

2.1 Overview



2.2 Gelation

1. Ensure that all components are allowed to equilibrate to room temperature before use and conduct all steps in a laminar flow hood at room temperature.
2. Dilute **Gel A**, **Gel B** and **Gel C** 1 in 5 with complete culture medium by adding directly to tubes containing gel concentrate (See kit contents on Page 3 for a dilution guide). Either mix by vortexing for 10 seconds or gently mix with a pipette until homogenous.
3. Carefully remove culture medium from each well of the plate.
4. Add 0.09 mL of the diluted (see step 1) **Gel A** Solution to each well.
5. Gently add 0.09 mL of the diluted (see step 1) **Gel B** Solution.
6. Add 0.05 mL of Gelation Buffer 1 (GB1) dropwise over the surface of Gel A/B solution. Allow **10 minutes** for Gelation.
7. Add 0.05 mL of Gelation Buffer 2 (GB2) dropwise over the surface of Gel A/B. Allow a further **10 minutes** for gelation.
8. Remove GB1/GB2 mixture from each well and wash for 5 minutes with 0.2 mL culture medium per well.
9. Remove culture medium and add 0.09 mL of the diluted (see step 1) **Gel C** solution to the centre of the gelled surface.

10. Place adhesive plate seal over the surface of the plate ensuring it is properly sealed.
11. Replace the lid and store away from light at an appropriate temperature.¹ Allow to settle for 4 hours before shipping.

¹Atelerix can recommend an appropriate storage temperature for your particular cell type. Please contact technical@atelerix.co.uk

2.3 Release

1. Ensure that all components are allowed to equilibrate to room temperature before use and conduct all steps in a laminar flow hood at room temperature.
2. Remove plate seals and infuse 0.1 mL of **Dissolution Buffer** using a P200 tip. Leave to de-gel for **5 minutes**. N.B. Pierce the surface of the gel with tip of pipette and infuse dissolution solution.
3. Remove 0.2 mL of well contents.
4. Add 0.2 mL of Dissolution Buffer and allow to de-gel for **10 minutes**.
5. Remove 0.4 mL (remaining contents of each well).
6. Replace each well with culture medium.
7. Return to normal culture conditions for **4 hours** or **overnight**.

2.4 Conditioning Chronos Advance CRT container

1. Remove the **six orange** 'Cool Phase' PCM panels from the CRT container and place them un-stacked at a temperature of approximately 20°C for **at least 24 hours**.
2. Reassemble the container placing the bottom and side PCM panels in place, ensuring that the side with the writing is facing towards outwards (as shown in the diagram below).
3. Place samples in the silver central payload box and carefully place in the centre of the CRT container.
4. Load the final PCM panel on top of the payload box ensuring that the side with the writing is facing outwards.
5. Place the polystyrene lid onto the system ensuring it is tightly sealed.
6. Close the outer carton.

Please consult Chronos Advance Pack-Out Instructions for a video detailing the assembly instructions.



3 Statements

3.1 Kit storage and stability

This kit is stable at 4°C for 6 months. Bring components up to room temperature before use.

Atelerix does not recommend using the kit after the expiry date stated on the packaging.

3.2 Trademarks

WellReady™ is a trademark of Atelerix Ltd.

3.3 Cellular material

Cellular monolayers can be used. Please ensure that biological material is free of fungal and bacteriological contamination before proceeding.

3.4 Trademarks

WellReady™ is a trademark of Atelerix Ltd.